


**PAKISTAN COUNCIL OF RESEARCH IN WATER RESOURCES  
MINISTRY OF WATER RESOURCES  
GOVERNMENT OF PAKISTAN**



**TENDER DOCUMENT**

**FOR**

**PROCUREMENT OF LABORATORY EQUIPMENT / INSTRUMENTS AND  
INSTALLTION OF SOLAR SYSTEM UNDER THE PROJECT “STRENGTHENING OF  
PCRWR WATER QUALITY LABORATORIES IN SINDH IN COLLABORATION  
WITH UNICEF”**

**Address:**

Pakistan Council of Research in Water Resources,  
Main University Road, Karachi.

Website: [www.pcrwr.gov.pk](http://www.pcrwr.gov.pk)

Email: [pcrwr.karachi@gmail.com](mailto:pcrwr.karachi@gmail.com)

Telephone: 021-34028062, 99333717-8

## **TABLE OF CONTENTS**

<b>Description</b>	<b>Page No.</b>
<b>INVITATION FOR BIDS</b>	<b>4</b>
<b>INSTRUCTIONS TO BIDDERS</b>	<b>6</b>
<b>A. General</b>	
IB.1 Scope of Bid	<b>7</b>
IB.2 Source of Funds	<b>7</b>
IB.3 Eligible Tenderers	<b>7</b>
IB.4 One Bid Per Bidder	<b>7</b>
IB.5 Cost of Bidding	<b>8</b>
<b>B. Bidding Documents</b>	
IB.7 Contents of Tendering Documents	<b>8</b>
IB.8 Clarification of Tendering Documents	<b>8</b>
IB.9 Amendment of Tendering Documents	<b>9</b>
<b>C. Preparation of Bids</b>	
IB.10. Language of Tender	<b>9</b>
IB.11 Documents Accompanying the Tender	<b>9</b>
IB.12 Tender Prices	<b>10</b>
IB.13 Currencies of Tender and Payment	<b>10</b>
IB.14 Tender Validity	<b>10</b>
IB.15 Tender Security	<b>10</b>
IB.18 Format and Signing of Bid	<b>11</b>
<b>D. Submission of Bids</b>	
IB.19 Sealing and Marking of Bids	<b>12</b>
IB.20 Deadline for Submission of Bids	<b>12</b>
IB.21 Late Bids	<b>12</b>
IB.22 Modification, Substitution and Withdrawal of Bids	<b>13</b>
<b>E. Bid Opening and Evaluation</b>	
IB.23 Bid Opening	<b>13</b>
IB.24 Process to be Confidential	<b>13</b>
IB.25 Clarification of Bids	<b>14</b>
IB.26 Examination of Bids and Determination of Responsiveness	<b>14</b>
IB.27 Correction of Errors	<b>14</b>

IB.28	Evaluation and Comparison of Bids	14
<b>F.</b>	<b>Award of Contract</b>	
IB.29	Award Criteria	16
IB.30	Employer's Right to Accept any Bid and to Reject any or all Bids	17
IB.31	Notification of Award	17
IB.32	Performance Security	17
IB.33	Signing of Contract Agreement	17
IB.34	General Performance of the Bidders	18
IB.35	Integrity Pact	18
IB.36	Instructions Not Part of Contract	18
	<b>TENDERING DATA</b>	<b>19</b>
	<b>FORMS</b>	
	FORM OF TENDER	34
	CONTRACT AGREEMENT	36
	INTEGRITY PACT	38

**INVITATION  
FOR  
BIDS**

## **INVITATION FOR BIDS**

### **(E-Bidding through EPADS)**

#### **National Competitive Bidding (Single Stage Two Envelope)**

Pakistan Council of Research in Water Resources (PCRWR) working under Ministry of Water Resources (MoWR), Government of Pakistan invites sealed bids from the firms/companies/ sole distributors/ suppliers registered with Federal Board of Revenue (FBR) (as per GST & Income Tax laws) and Sindh Revenue Board (SRB) for particulars given below through EPADS:

<b>Sr.#</b>	<b>Description</b>	<b>Estimated Cost (PKR)</b>
1.	Procurement (Supply, Installation) of Laboratory Equipment & Instruments at WQL Hyderabad, Tandojam and Nawabshah	33 Million Pakistani Rupees
2.	Capacity Building of Staff on the Equipment/Instruments at WQL Hyderabad, Tandojam and Nawabshah	
3.	Supply, Installation of Solar System at WQL Tandojam and Nawabshah	05 Million Pakistani Rupees
4.	Estimated Total cost of Bid (With all taxes)	38 Million Pakistani Rupees
5.	Last date of Submission	<b>25.08.2025 (11:00 AM)</b>
6.	Date and time of opening of Bids	<b>25.08.2025 (11:30 AM)</b>
7.	Bid/tender Security	<b>5% (PKR)</b>

1. Regional/Project Director, PCRWR Regional Office Karachi, invites E-bids through EProcurement (EPADS) under PPRA Rules, for the above mentioned scope of work from eligible bidders.
2. Bidding shall be conducted through E-Procurement (Open Competitive Bidding Single Stage Two Envelop) procedure and is open to all eligible bidders as defined in the bidding document.
3. Interested and registered eligible bidders on E-PADS may obtain bidding documents from EPADS, PPRA and Departmental Website respectively.
4. Interested and registered eligible bidder can participate through E-Procurement (EPADS) (<http://eprocare.gov.pk/>).
5. E-bids must be submitted through E-Procurement (EPADS) on or before **25<sup>th</sup> August, 2025** till 11:00 am. Proposals must be accompanied by a Bid Security (5 % of Total Quoted Amount) in the form of Demand Draft, Pay Order or Call at Deposit duly issued from a Scheduled Bank of Pakistan.
6. Hard copy of Documents must be submitted physically at Committee Room of Pakistan Council of Research in Water Resources Main University Road Karachi.
7. E-bids will be opened on **25<sup>th</sup> August 2025** at **11:30 am** in the presence of bidders or their representatives. If bid closing/opening date falls on local/national holiday, the date of bid closing/opening shall be the next working day on the same time and venue.
8. The bidders are required to give their best and final prices (inclusive of all applicable Taxes) as no negotiations are expected.
9. Bids not complying with the requirements shall be rejected being nonresponsive. Alternate proposals and / or conditional bids shall be considered non-responsive.
10. Taxes will be deducted as per applicable government rules. Bidders must have Valid NTN and Sales Tax registration certificates.
11. Bids which are incomplete, not E-bids through EPADS, not signed and / or not stamped, late or submitted by other than specified mode will be rejected.

12. The contract shall be awarded to the bidder who meets all the technical evaluation criteria and is either the most technically qualified or the lowest responsive bidder fulfilling the required specifications.
13. This advertisement and bidding documents are also available on E-PADS, PPRA website and PCRWR website [www.pcrwr.gov.pk](http://www.pcrwr.gov.pk)]

Dr. Ghulam Murtaza  
Regional Director/Project Director  
021-34028062  
pcrwr.karachi@gmail.com

**INSTRUCTIONS  
TO  
BIDDERS**

## **INSTRUCTIONS TO BIDDERS**

(Note: These Instructions along with Bidding Data will not be part of Contract and will cease to have effect once the Contract is signed).

### **A. GENERAL**

#### **IB.1 Scope of Tender**

##### **1.1 Scope of Tender**

The Regional Director / Project Director, PCRWR (hereinafter referred to as “**the Employer**”), as defined in the Tendering Data, invites sealed bids for the Procurement of Services for the Installation of Multi-Level Observation Wells (MLOWs) in province of Sindh, as outlined in the Tendering Data. The firms or individuals submitting bids shall hereinafter be referred to as “**the Tenderers.**”

Tenderers should quote for the work specified with its unit rate and its total rate with all Applicable Taxes.

The bids must be quoted in Pak Rupees and submitted to the undersigned by **25<sup>th</sup> August, 2025 at 11:00 Hrs.**

The tenders will be opened on the same date at 1130 hrs.

#### **IB.2 Source of Funds**

- 2.1 The PCRWR have arranged funds from UNICEF funded project “Strengthening PCRWR Water Quality Laboratories in Sindh under SDG Aligned Monitoring & Management in Collaboration with UNICEF”

#### **IB.3 Eligible Tenderers**

- 3.1 Tendering is open to all eligible firms/companies/sole distributors/ suppliers which have active registration with the Federal Board of Revenue (FBR)/SRB. Eligible bidders must demonstrate adequate financial and technical capacity, along with relevant experience, to undertake the Procurement of Laboratory Equipment /Instruments. Preference will be given to firms with demonstrated experience in supplying equipment/instruments to water testing laboratories, conducting capacity building of laboratory technicians, and/or those acting as sole distributors of the manufacturers.

#### **IB.4 One Bid per Bidder**

- 4.1 Each bidder shall submit only one bid either by himself, or as a partner in a joint venture. A bidder who submits or participates in more than one bid (other than

alternatives pursuant to Clause IB.16) will be disqualified.

### **IB.5 Cost of Bidding/Tendering**

- 5.1 The tenderer shall bear all costs associated with the preparation and submission of its tender and PCRWR will in no case be responsible or liable for those costs, regardless of the conduct or outcome of the tendering process

## **B. TENDER DOCUMENTS**

### **IB.7 Contents of Tender Documents**

- 7.1 The Bidding Documents, in addition to invitation for bids, are those stated below and should be read in conjunction with any Addenda issued in accordance with Clause IB.9.

1. Instructions to Tenderers & Tendering Data
2. Forms of Tender
3. Standard Forms:
  - (i) Form of Performance Security
  - (ii) Form of Contract Agreement
4. Integrity Pact
5. Specification

- 7.2 The bidders are expected to examine carefully the contents of all the above documents. Failure to comply with the requirements of bid submission will be at the Bidder's own risk. Pursuant to Clause IB.26, bids which are not substantially responsive to the requirements of the Bidding Documents will be rejected.

### **IT.8 Clarification of Tender Documents**

- 8.1 A prospective tenderer requiring any clarification(s) in respect of the Tender Documents may notify the Regional Director, PCRWR against subject tender **“Strengthening PCRWR Water Quality Laboratories in Sindh under SDG Aligned Monitoring & Management in Collaboration with UNICEF”** at the address indicated in the Tendering Data.
- 8.2 The Regional Director, PCRWR will respond to any request for clarification which it receives earlier than 10 days prior to the deadline for the submission of tenders. Copies of the response will be forwarded to all prospective tenderers, at least 5 days prior to dead line for submission of Tenders, who have received the Tender Documents including a description of the enquiry but without identifying its source.

## **IB.9 Amendment of Tender Documents**

- 9.1 At any time prior to the deadline for submission of tenders, the Regional / Project Director, PCRWR, Karachi may for any reason, whether at his own initiative or in response to a clarification requested by a respective tenderer, modify the Tender Documents by issuing addendum.
- 9.2 Any addendum thus issued shall be part of the Tender Documents pursuant to Sub-Clause 7.1 hereof, and shall be communicated in writing to all purchasers of the Tender Documents. Prospective tenderers shall acknowledge receipt of each addendum in writing to the Regional /Project Director, PCRWR, Karachi
- 9.3 To afford prospective tenderers reasonable time in which to take an addendum into account in preparing their tenders, the Regional Director /Project Director, PCRWR may extend the deadline for submission of tenders. Clause IB.20

## **C. PREPARATION OF TENDERS**

### **IB.10 Language of Tender**

- 10.1 The tender prepared by the tender and all correspondence and documents relating to the Tender, exchanged by the tenderer and the Regional Director /Project Director, PCRWR, shall be written in the English language, provided that any printed literature furnished by the tenderer may be written in another language so long as accompanied by an English translation of its pertinent passages in which case, for purposes of interpretation of the Tender, the English translation shall govern.

### **IB.11 Documents Comprising the Tender**

- 11.1 The tender prepared by the tenderer shall comprise the following components:
  - (a) Covering Letter
  - (b) Form of Tender duly filled, signed and sealed, in accordance with Clause IB.18.3.
  - (c) Tender Security furnished in accordance with Clause IB.15.
  - (d) Power of Attorney in accordance with Clause IB 18.5.
  - (e) Documentary evidence in accordance with Clause IB.11
  - (f) Documentary evidence in accordance with Clause IB.I2.
- 11.2 Pursuant to Clause IB.11, the tenderer shall furnish, as part of its tender, documents establishing the tender's eligibility to tender and its qualifications to perform the Contract if its tender is accepted.
- 11.3 Tenderer must possess and provide evidence of the experience as stipulated in Tendering Data.

- 11.4 The documentary evidence of the Works' conformity to the Tender Documents may be in the form of literature, drawings and data and shall furnish documentation as set out in Tendering Data.

#### **IB.12 Tender Prices**

- 12.1 Unless stated otherwise in the Bidding Documents, the Contract shall be for the whole of the Works as described in Sub-Clause 1.1 hereof, based on the unit rates and / or prices submitted by the bidder.
- 12.2 The bidders shall fill in rates and prices for all items of the supplies described in the tender document.
- 12.3 Unless otherwise stipulated in the Conditions of Contract, prices quoted by the tenderer shall remain fixed during the tenderer's performance of the Contract and not subject to variation on any account.
- 12.4 The tenderer shall note that standards for material and equipment, and references to brand names or catalogue numbers, designated by the Employer in the Technical Provisions are intended to be descriptive only and not restrictive.

#### **IB.13 Currencies of Bid and Payment**

- 13.1 The unit rates and prices in the Schedule of Prices shall be quoted by the tenderer in the currency as stipulated in Tendering Data.
- 13.2 **The tenderer will provide the total cost indicating all applicable taxes.**

#### **IB.14 Tender Validity**

- 14.1 Tenders shall remain valid for the period of **210 days** as stipulated in the Tendering Data after the date of tender opening.

#### **IB.15 Tender Security**

- 15.1 Each tenderer shall furnish, as part of his tender, a Tender Security of **5%** of the total bid price in Pak. Rupees in the form of a Bank Guarantee issued by a Scheduled Bank in Pakistan in favour of the Employer valid for a period **28 days** beyond the tender validity date.
- 15.2 Any tender not accompanied by an acceptable Tender Security shall be rejected by the Employer as non-responsive.
- 15.3 The tender securities of unsuccessful tenderers will be returned upon award of contract to the successful tenderer or on the expiry of validity of Tender Security whichever is earlier.

- 15.4 The Tender Security of the successful tenderer will be returned when the tenderer has furnished the required Performance Security, pursuant to Clause IB.32 and signed the Contract Agreement, pursuant to Clause IB.33.1.
- 15.5 The Tender Security may be forfeited:
- (a) if a tenderer withdraws his tender during the period of tender validity;
  - (b) if a tenderer does not accept the correction of his Tender Price, pursuant to Clause IB 27.1 hereof; or
  - (c) in the case of a successful tenderer, if he fail to:
    - (i) Furnish the required Performance Security in accordance with Clause IB.32, or
    - (ii) Sign the Contract Agreement, in accordance with Clause IB.33.1

### **IB.18 Format and Signing of Bid**

- 18.1 Bidders are particularly directed that the amount entered on the Form of Bid shall be for performing the Contract strictly in accordance with the Bidding Documents.
- 18.2 All appendices to Bid are to be properly completed and signed.
- 18.3 No alteration is to be made in the Form of Bid nor in the Appendices thereto except in filling up the blanks as directed. If any such alterations be made or if these instructions be not fully complied with, the bid may be rejected.
- 18.4 Each bidder shall prepare one (1) original and the number of copies specified in the Bidding Data of the complete set of bid documents, as described in Clause IB.7, without any alterations. Each set shall be clearly marked as '**ORIGINAL**' or '**COPY**' as appropriate. In the event of any discrepancy between the original and the copies, the original shall prevail. The '**Technical Bid**' and the '**Financial Bid**' must be submitted separately."
- 18.5 The original and all copies of the bid shall be typed or written in indelible ink (in the case of copies, Photostats are also acceptable) and shall be signed by a person or persons duly authorized to sign on behalf of the bidder pursuant to Sub- Clause 11.1(a) hereof. **All pages of the bid shall be initialed and stamped by the person or persons signing the bid.**
- 18.6 The bid shall contain no alterations, omissions or additions, except to comply with instructions issued by the Regional Director/Project Director, PCRWR or as are necessary to correct errors made by the bidder, in which case such corrections shall be initialed by the person or persons signing the bid.
- 18.8 Bidders should retain a copy of the Bidding Documents as their file copy.

## **D-SUBMISSION OF TENDER**

### **IB.19 Sealing and Marking of Bids**

- 19.1 Each bidder shall submit his bid as through EPADS as under:
- (a) ORIGINAL and each copy of the Bid shall be separately sealed and put in separate envelopes and marked as such.
  - (b) The envelopes containing the 'ORIGINAL' and 'COPIES' will be put in one sealed envelope and addressed / identified as given in Sub- Clause 19.2 hereof for both 'TECHNICAL' as well 'FINANCIAL' Bid.
- 19.2 The inner and outer envelopes shall:
- (a) Be addressed to the Employer at the address provided in the Bidding Data
  - (b) Bear the name and identification number of the contract as defined in the Bidding Data; and
  - (c) Provide a warning not to open before the time and date for bid opening, as specified in the Bidding Data.
- 19.3 In addition to the identification required in Sub- Clause 19.2 hereof, the inner envelope shall indicate the name and address of the bidder to enable the bid to be returned unopened in case it is declared "late" pursuant to Clause IB.21
- 19.4 If the outer envelope is not sealed and marked as above, the Employer will assume no responsibility for the misplacement or premature opening of the Bid.

### **IB.20 Deadline for Submission of Bids**

- 20.1 Bids must be received by the Employer no later than the time and date stipulated in the Bidding Data.
- 20.2 The Employer may, at his discretion, extend the deadline for submission of bids by issuing an amendment in accordance with Clause IB.9, in which case all rights and obligations of the Employer and the bidders previously subject to the original deadline will thereafter be subject to the deadline as extended.

### **IB.21 Late Bids**

- 21.1 (a) Any bid received by the Employer after the deadline for submission of bids prescribed in Clause IB.20 will be returned unopened to such bidder or will be rejected.

## **IB.22 Modification, Substitution and Withdrawal of Bids**

- 22.1 Any bidder may modify, substitute or withdraw his bid after bid submission provided that the modification, substitution or written notice of withdrawal is received by the Employer prior to the deadline for submission of bids.
- 22.2 The modification, substitution, or notice for withdrawal of any bid shall be prepared, sealed, marked and delivered in accordance with the provisions of Clause IB.19 with the outer and inner envelopes additionally marked “MODIFICATION”, “SUBSTITUTION” or “WITHDRAWAL” as appropriate.
- 22.3 No bid may be modified by a bidder after the deadline for submission of bids except in accordance with Sub-Clauses 22.1 and 27.2.
- 22.4 Withdrawal of a bid during the interval between the deadline for submission of bids and the expiration of the period of bid validity specified in the Form of Bid may result in forfeiture of the Bid Security in pursuance to Clause IB.15.

## **E. TENDER OPENING AND EVALUATION**

### **IB.23 Tender Opening**

- 23.1 The Regional Director/Project Director, PCRWR will open the tenders, in the presence of tenderers' representatives who choose to attend, at the time, date and location stipulated in the Tendering Data.
- 23.2 (a) The tenderer's name, Tender Prices, any discount, the presence or absence of Tender Security, and such other details as the Employer at its discretion may consider appropriate, will be announced by the Regional Director /Project Director, PCRWR at the time of tender opening.  
  
(b) Any Tender Price or discount which is not read out and recorded at tender opening will not be taken into account in the evaluation of tender.

### **IB.24 Process to be Confidential**

- 24.1 Subject to Clause IB 25.1 heretofore, no tenderer shall contact the PCRWR on any matter relating to its Tender from the time of the Tender opening to the time the Contract is awarded.
- 24.2 Any effort by a tenderer to influence Regional Director/Project Director PCRWR in the Tender evaluation, Tender comparison or Contract Award decisions may result in the rejection of his Tender.

## **IB.25 Clarification of Bids**

- 25.1 To assist in the examination, evaluation and comparison of Tenders the Regional Director/Project Director, PCRWR may in its discretion ask the tenderer for a clarification of its Tender. The request for clarification and the response shall be in writing and no change in the price or substance of the Tender shall be sought, offered or permitted.

## **IB.26 Examination of Bids and Determination of Responsiveness**

- 26.1 (a) Prior to the detailed evaluation, the Regional Director/Project Director, PCRWR will determine the substantial responsiveness of each tender to the Tender Documents. For purpose of these Clauses, a substantially responsive tender is one which conforms to all the terms and conditions of the Tender Documents without material deviations. It will include determining the requirements listed in Tendering data.

## **IB.27 Correction of Errors**

- 27.1 Arithmetical errors will be rectified on the following basis:

If there is a discrepancy between the unit price and total price that is obtained by multiplying the unit price and quantity, the unit price shall prevail and the total price shall be corrected. If there is a discrepancy between the words and figures the amount in words shall prevail. If there is a discrepancy between the Total Tender price entered in Form of Tender and the total shown in Schedule of Prices-Summary, the amount stated in the Form of Tender will be corrected by the PCRWR in accordance with the Corrected Schedule of Prices.

If the tenderer does not accept the corrected amount of Tender, his Tender will be rejected and his Tender Security forfeited.

- 27.2. A Tender determined as substantially non-responsive will be rejected and will not subsequently be made responsive by the tenderer by correction of the non-conformity.
- 27.3 Any minor informality or non-conformity or irregularity in a Tender which does not constitute a material deviation may be waived by the Regional Director /Project Director, PCRWR provided such waiver does not prejudice or affect the relative ranking of any other tenderers.

## **IB.28 Evaluation and Comparison of Bids**

- 28.1 The Tender Evaluation Committee PCRWR will evaluate and compare only the tenders previously determined to be substantially responsive pursuant to Sub-clauses IB 26.1 and 27.2 to 27.3 as per requirements given hereunder. Tenders will be evaluated for complete scope of works. The prices will be compared on the basis

of the Evaluated Tender Price pursuant to Sub-Clause 28.2 herein below.

(a) Technical Evaluation

It will be examined in detail whether the items offered by the tenderer complies with the Technical Provisions of the Tender Documents. For this purpose, the tenderer's data submitted will be compared with technical features/criteria of the Tendering Data detailed in the Technical Provisions. Other technical information submitted with the tender regarding the Tendering Data will also be reviewed.

(b) Commercial Evaluation

It will be examined in detail whether the tenders comply with the commercial/contractual conditions of the Tender Documents. It is expected that no major deviation/stipulation shall be taken by the tenderers.

## **28.2 Evaluated Tender Price**

In evaluating the tenders, the Regional Director/ Project Director, PCRWR will determine for each tender in addition to the Tender Price, the following factors (adjustments) in the manner and to the extent indicated below to determine the Evaluated Tender Price:

- (i) Making any correction for errors pursuant to Sub-Clause 26.1 & 27.1 hereof.
- (ii) making an appropriate price adjustment for any other acceptable variation or deviation.
- (iii) Making an appropriate price adjustment for Deviations in terms of Payments (if any and acceptable to Tender Evaluation Committee PCRWR).

## **28.3 Evaluation Methods**

Pursuant to Sub-Clause 28.2, Para (ii), and (iii) following evaluation methods will be followed:

**a) Technical soundness: 60% weightage**

Soundness of firm / experience: 40 marks

Quoted specifications matching with the tender specifications: 60 marks

**b) Financial soundness: 40% weightage**

Lowest bidder: 100 marks

Second lowest bidder: 90 marks

Third lowest bidder: 80 marks

Other lowest bidders will be ranked with 10 less marks in descending order.

The tenderer/ bidder will provide past performance /Experience of the **last 3 years**

along with supporting documents (supply orders and performance/completion certificates).

(i) Price Adjustment for Technical Compliance

The cost of making good any deficiency resulting from technical noncompliance will be added to the Corrected Total Tender Price for comparison purposes only. The adjustments will be applied taking the highest price quoted by other tenderers being evaluated in detail in their original Tenders for corresponding item. In case of non-availability of price from other tenderers, the price will be estimated by Project Director.

(ii) Price Adjustment for Commercial Compliance

The cost of making good any deficiency resulting from any quantifiable variations and deviations from the Tender Schedules and Conditions of Contract, as determined by Project Director will be added to the Corrected Total Tender Price for comparison purpose only. Adjustment for commercial compliance will be added to the Corrected Total Tender Prices.

(iii) Price Adjustment for Deviation in Terms: If Payments Refer to Tendering Data

(iv) **Guarantee/Warranty**

The warrantee period is **six months** which should be provided on proper legally recognized document. In case of failure, the same will be arranged from the performance security amount. It will be the responsibility of supplier to arrange any spare part, or repair in case of damage, for the period of one year.

The representative of PCRWR may examine any item, equipment, material, supplies and/or services to be provided under this contract, at any time prior to expiry of this contract. He may reject or require alteration thereto or replacements thereof as a condition of acceptance and its findings shall be conclusive except in regard to defects or fraud which may appear later.

## **F. AWARD OF CONTRACT**

### **IB.29 Award Criteria**

- 29.1 Subject to Sub-clause IB 30 and IB.34, the Regional Director/Project Director, PCRWR will award the Contract to the tenderer whose tender has been determined to be substantially responsive to the Tender Documents fulfilling all technical specification and who has offered the lowest evaluated Tender Price, provided that such tenderer has been determined to be qualified to satisfactory perform the

Contract in accordance with the provisions of Clause IB.3 and qualify pursuant to Sub-Clause IB 29.2.

- 29.2 In the absence of pre-qualification, the Regional Director/Project Director, PCRWR will determine to its satisfaction whether the substantially responsive, fulfilling technical requirements and is the lowest evaluated tenderer is qualified to satisfactorily perform the Contract.
- 29.3 The determination will take into account the tenderer's financial and technical capabilities. It will be based upon an examination of the documentary evidence of the tenderer's qualifications submitted under Clause IB.11, as well as such other information as the Regional Director/ Project Director PCRWR deems necessary and appropriate

### **IB.30 Employer's Right to accept any Bid and to reject any or all Bids**

- 30.1 Notwithstanding Sub-clause IB.29 the Regional Director/ Project Director PCRWR reserves the right to accept or reject any tender, and to annul the tendering process and reject all tenders, at any time prior to award of Contract, without thereby incurring any liability to the affected tenderers or any obligation to inform the affected tenderers of the grounds for the PCRWR action.

### **IB.31 Notification of Award**

- 31.1 Prior to expiration of the period of tender validity prescribed by the PCRWR, the PCRWR will notify the successful tenderer in writing ("Letter of Acceptance") that *his/her* tender has been accepted.

### **IB.32 Performance Security**

- 32.1 The successful tenderer shall furnish to the PCRWR a Performance Security in the form and the amount stipulated in the Conditions of Contract within a period of 14 days after the receipt of Acceptance.
- 32.2 Failure of the successful tenderer to comply with the requirements of Sub-clauses IB 32.1 & IB 33 or clause IB.35 shall constitute sufficient grounds for the annulment of the award and forfeiture of the Tender Security.

### **IB.33 Signing of Contract Agreement**

- 33.1 Within 14 days from the date of furnishing of acceptable Performance Security under the Conditions of Contract, the PCRWR will send the successful tenderer the Form of Contract Agreement provided in the Tender Documents, incorporating all agreements between the parties.

**IB.34 General Performance of the Bidders**

The Employer reserves the right to obtain information regarding the performance of bidders on previously awarded contracts/works. In cases of consistently poor performance, as reported by the respective employers of those contracts, the Employer may, inter alia, reject the bid of such a bidder.

**IB.35 Integrity Pact**

The Bidder shall sign and stamp the Integrity Pact provided to Bid in the Bidding Documents for all Federal Government procurement contracts exceeding Rupees ten million. Failure to provide such Integrity Pact shall make the bidder non-responsive.

**IB.36 Instructions not Part of Contract**

Bids shall be prepared and submitted in accordance with these Instructions which are provided to assist bidders in preparing their bids, and do not constitute part of the Bid or the Contract Documents.

# **TENDERING DATA**

## Tendering Data

### **SCOPE OF WORK:**

The selected firm shall be responsible for the following:

1. Procurement and Supply of laboratory equipment/instruments, including all necessary accessories, parts, and standards.
2. Installation of the supplied equipment/instruments at designated Water Quality Laboratories located in Hyderabad, Tandojam, and Nawabshah.
3. Capacity Building and Training of laboratory staff on the proper use, operation, and maintenance of the supplied equipment/instruments.
4. Provision of a Six Months Service Warranty for all supplied equipment/instruments, covering maintenance and technical support.
5. Supply and Installation of a Solar Power System at the Water Quality Laboratories, complete in all respects and fully functional upon commissioning.
6. **Completion Timeline:** All activities shall be completed within 30 days from the date of issuance of the official Supply/Work Order.
7. Please note that the quantities of Items in BOQ can be increased or decreased subject to availability of Funds.

The tender document contains items with specifications as under:

**(A) LAB EQUIPMENT/INSTRUMENTS:**

Sr. No.	Description	Unit	Unit Rate (PKR) With Taxes	Amount (PKR) With Taxes
1	<p><b>Multiparameters Colorimeter (Spectrophotometer)</b></p> <p><b>Features:</b>            Operating Mode: Transmittance (%), absorbance and concentration (wavelength, time),            Source Lamp: Tungsten (visible range), deuterium (UV range),            Wavelength Range: 190- 1100 nm,            Wavelength Accuracy: <math>\pm 1</math>nm,            Wavelength Reproducibility: <math>&lt; 0.1</math> nm,            Wavelength Resolution 0.1 nm,            Spectral Bandwidth: 2nm,            Spectral scope: 4nm            Photometric range: -3.3 to +3.3Abs            Photometric Resolution: Abs: 0, 001 Transmission: 0,1%            Scanning Speed: 900 nm/min (in 1 nm steps),            Programmed methods: <math>&gt;250</math>            Standards (solutions/ powder pillows) for NO<sub>3</sub>, NO<sub>2</sub>, Fe, SO<sub>4</sub>, PO<sub>4</sub>, Fluoride (10 for each) 1 unit=100 test            Additional 2 pair of Cuvettes            All other related accessories, parts, Automatic pipettes 1- pml with steeples volume adjustment (digital), pipette 1- 5ml (pack with 100pc), Automatic pipette 0.1-1ml with steeples volume adjustment (digital), secondary standards for VIS, UV-VIS with DAkkS Calibration, software, manuals.            OR Equivalent.</p>	03		
2	<p><b>Flame Photometer:</b></p> <p><b>Features:</b>            Digital Flame Photometer Microprocessor Controlled, Touch screen.            Elements: Na, K, Ca, Ba            Reproducibility: <math>\leq 2\%</math>,            Response Time: <math>&lt; 8</math> seconds,            Sample Uptake Volume (ml): <math>&lt; 6</math>ml/minute,            Air compressor: 1450 – 1700 rpm (Single Cylinder Piston Compressor with air tank – OIL FREE), Adjustable Pressure Range = 0-6 Bar, Air Tank =3 liters.            Gas cylinder (10-12 kg capacity minimum)            Fuel: Natural Gas or LPG            Gas regulator            Standards 1000 ppm (Na, K, Ba) 500ml x 4 for Na &amp;K and 500ml x 2 for Ba.</p>	03		

	All other related accessories, parts, software, manuals etc OR Equivalent.			
3	<p><b>Multipara meter Instrument (pH/ Redox,DO, Cond./TDS):</b></p> <p><b>Features:</b>  pH= -2.0 to 16.0  Redox/ORP: -200 TO +2000Mv  Temp. -10.0 to +150 °C  Accuracy: pH ± 0.25%FS  Redox/ORP: 0.25%FS  Temp. 0.25%FS  Calibration: 1-5-point calibration automatic and or manual  Operational Conditions: Temperature -25 °C to +150 °C, Humidity 0 to 95%, storage temp. -25 °C to +70 °C  Data storage: internal mass storage 8-10 GB, Protection Glass: IP 67. Power supply: NiMH batteries of via USB-interface.  OR Equivalent.</p>	03		
4	<p><b>Turbidity Meter:</b></p> <p><b>Features:</b>  Measuring range: 0 to 4000 NTU  Principle: Nephelometric  Resolution (max): 0.01  Accuracy: ± 1.8%  Stray light: &lt;0.014 NTU  Display: Touch screen  Data Transfer: USB Interface  Data Logger: 250 Measurements (minimum), all calibrations, verifications and factory re-store functions  Complete set, along with sample cell with lids, calibration standards, software, manuals, power batteries.  One- Additional set of calibration standards and sample cell.  OR Equivalent</p>	03		
5	<p><b>Color Measurement (E-Comparator):</b></p> <p><b>Features:</b>  Light Source: White LED  Sensors: Tristimulus Detectors (Reference &amp; Cell)  Color: Pt-Co  Measuring Range: 2-500  Resolution: 1 Pt-Co Unit  Repeatability: ± 3% +1 Pt-Co Unit  Detection Limit: 2 Pt-Co TCU Unit  Data storage: ≥ 20,000 readings  Temp.: 80 °C (sample Temp Max)  Comparator View: 2 Field</p>	03		

	<p>Touch screen: Resistiv</p> <p>Along with Software, Optical glass cell, plastic cell, liquid reference standards for zero, 1, 3,5, 10,15,30,50 &amp; 100 Pt-Co. one additional set of Optical Glass and plastic cell and all other accessories and parts, manuals.</p> <p>OR Equivalent.</p>			
6	<p><b>Photometer for COD with Digester/Reactor:</b></p> <p><b>Features:</b></p> <p><b>Photometer:</b> with high quality log-life LED paired with interference filters.</p> <p>Filter wavelengths: 430,530,560,580,610,660nm</p> <p>Accuracy: <math>\pm 1</math>nm</p> <p>Bandwidth: 5-6nm</p> <p>Photometric accuracy: <math>\pm 3\%</math></p> <p>Resolution: 0.01A</p> <p>Operating Environment: 5-40 °C</p> <p>Alongwith all accessories, batteries, manual etc</p> <p>With COD vial range 3-150mg/l, 15-300mg/l, 20-1500mg/l, 200-15000mg/l (50 pc each)</p> <p><b>Thermoreactor (Tube test Digestion):</b></p> <p>Cuvette recording :24 slots</p> <p>Temp. Supervision: Pt 100 A class</p> <p>Temperature selectable 100/120/150 °C</p> <p>Temp. stability: <math>\pm 1</math> °C</p> <p>Selected time: 30/60/120min and continuous operation</p> <p>Protection against overheating: at the Alu block at 190 °C</p> <p>Protection grid: PPS</p> <p>Lid: PPS</p> <p>Heating Block: Aluminum</p> <p>Block insert: PBT</p> <p>OR Equivalent.</p>	03		

7	<p><b>Hot Air Oven:</b></p> <p><b>Features:</b>  Volume: 108 L (3.8 ft<sup>3</sup>)  Temp Range (set): +20 °C – +300 °C  Working Range: Ambient +5 °C to +300 °C  Convection Mode: Natural convection (no fan)  Air Flap Adjustment: Electronically in 10 % increments  Temperature Resolution: 0.1 °C (≤99.9 °C), 0.5 °C (&gt;99.9 °C)  Controller: PID-microprocessor with TFT SingleDISPLAY  Timer: 1 min to 99 days; SetpointWAIT  Shelves: 2 included; up to 5 max  Max Load per Shelf: 20 kg  Max Chamber Load: 175 kg  Safety Systems: TB limiter, AutoSAFETY, alarms  Data Logging: ≥ 10-year capacity; AtmoCONTROL compatible  Electrical Ratings: 230 V / 2,800 W or 115 V / 1,800 W  Chamber Dimensions: 560 × 480 × 400 mm  Exterior Dimensions: 745 × 864 × 584 mm (+door handle)  Certifications: CE, VDE, UL, EAC  Calibration Certificate: at +160 °C OR Equivalent.</p>	02		
8	<p><b>Spark Free Laboratory Cabinet:</b></p> <p>Capacity: ~220 L  Temp Range: +1 °C to +15 °C (adjustable)  Internal Dimensions: ~470 × 440 × 1,062 mm  External Dimensions: ~600 × 610 × 1,250 mm  Shelves: 4 glass shelves (adjustable)  Cooling System: Compressor + fan; "fan-stop" feature  Interior/Exterior: PS plastic / steel housing with coating  Door Features: Lockable, convertible hinge, magnetic seal  Power: 220–240 V, ~0.8 kWh/24 h  Noise Level: ~47 dB sound power  Compliant With: BG-I 850-0, DGUV 213-850, CE, EU directives  OR Equivalent.</p>	04		
9	<p><b>Dissolved Oxygen Meter:</b></p> <p>DO Range: 0.1 – 20.0 mg/L / 1–200 % saturation  Resolution: 0.1 mg/L; 1 % saturation  Accuracy: ±0.1 mg/L (0–8 mg/L); ±0.2 mg/L (8–20 mg/L)  Probe: Intellical LDO10101 (1 m) or LDO10105 (5 m rugged)  Compensation: Automatic temp &amp; pressure; salinity manual entry  Display: Mono-TFT; up to 3 parameters at once  Data Memory: Up to 5,000 data points; GLP-compliant  IP Rating: IP67 (with battery cover installed)  Power: Rechargeable Li-ion; USB adapter 100–240 VAC  Operating Conditions: 0–60 °C ambient; &lt;90 % (non-condensing) humidity  Interface: Soft-touch keypad; multiple read modes</p>	03		

	Warranty:12 months standard. OR Equivalent.			
10	<b>Membrane Filtration Assembly:</b> Filtration Cup / Funnel: 300 mL borosilicate glass funnel (47 mm) with stainless/PTFE support disc Receiver Flask: 1 L GL-45 borosilicate glass flask with silicone stopper Vacuum Pump: Oil-free diaphragm or piston pump; ~70–100 mbar vacuum; low noise; thermal protection Flow Control: Manual vacuum regulator valve with pressure gauge Tubing & Connectors: Silicone tubing and adaptors to link funnel to pump Optional Safety Parts: Moisture trap or inline filter to protect pump OR Equivalent.	02		
11	<b>pH meter:</b> pH Range: 0.00 to 14.00 pH Resolution: 0.01 pH Accuracy: ±0.07 pH (pH 5–9); ±0.1 pH (pH 4–10); ±0.2 pH (outside range) Temperature Operating: 0–50 °C Humidity Operating: Up to 80 % RH (non-condensing) Power Supply: 9 V battery (006P) Display: Two-decimal LCD, digit height ~21.5 mm Protection: Water-resistant panel; rugged casing with electrode holder Connector: BNC standard Electrode Temp Range: –5 °C to +80 °C Along with pH standard pH 4, pH7, pH 10 (500ml each) OR Equivalent.	03		
12	<b>Electric Conductivity /TDS meter:</b> Conductivity Range: 0–200 mS/cm (auto-ranging) Resolution: 0.1 mS, depending on range Accuracy: ±0.5 % reading ±0.5 % full-scale Temperature Compensation: Automatic (NTC), linear or non-linear per EN 27888 Calibration: Single-point (1000–2000 µS/cm), adjustable cell constant Memory Capacity: 20 data sets with temp & optional location labeling Protection & Housing: IP67 waterproof, shock-resistant armor, built-in electrode holder Power: 4 × AA batteries; up to 1,500 h; auto-off Operating Temp: Meter: 0–50 °C; Cell: –5–80 °C Sensor Type: 4-pole graphite cell, integrated NTC sensor Display Functions: Conductivity, TDS, salinity, temperature display; Min/Max and Hold features OR Equivalent.	03		

13	<b>Refrigerator for Lab.</b> Laboratory Refrigerator, Inverter 18-20cft, Made Local, Dawlance or Equivalent	2		
<b>Total with All Taxes (PKR)</b>				

**(B) SUPPLY AND INSTALLATION OF SOLAR SYSTEMS**

Sr. No.	Description	Unit	Unit Rate (PKR) With Taxes	Amount (PKR) With Taxes
1	<p><b>Solar Panel 585 watts (12 panels 7020 watt):</b> 585 Watt Poly or Mono Crystalline Solar Panel (12-Panel Array (~7 kW Total)), Efficiency: ~22.65 - ~23 %, Temp. range: -40 °C to +85 °C, system voltage compatibility up to 1500 V, A Grade, Tier 01, Make China, Brand JINCO, LONGI, Canadian Solar or equivalent</p> <p><b>Inverter 6 KVA Hybrid 48 V:</b> Hybrid, MPPT, 48 V, 6KVA, Maximum Efficiency: ~97.6% MPPT Voltage Range: ~150 V to 425 V (start at ~125 V; allows high DC voltage design flexibility) Max PV Charge Current: ~135 A total (e.g. 2x~67 A per MPPT) PV Input Current: ~13 A + 13 A per MPPT (approximate) Transfer Time: 4 ms (seamless switching between solar, battery, and grid) Noise Level: &lt;30 dB (quieter operation) Make China, Brand Inverex "Nitrox" series or equivalent</p> <p><b>Lithium Batteries 100 AMP (100 AMP x2):</b> Rated Voltage:48 V (15S) Capacity / Energy: 100 Ah / 4.8 kWh Max Charge/Discharge: 100 A (1 C) Recommended Charge: ~20 A (0.2 C) Cut-off Voltage: ~40.5 V Charge Voltage: 54.0 ± 0.5 V Weight: ~38-42 kg Form Factor: 3U rack mount, 442 × 400 × 133 mm Operating Temp Range: Discharge -20 °C to +60 °C; Charge 0 °C to +60 °C Parallel Capability: Up to 8 modules (16 at reduced discharge rate) Communication Interfaces: RS232, 2 × RS485, SOC/ALM LEDs, optional LCD BMS Protections Included: Over-/under-voltage, over-/under-temperature, overcurrent, short circuit Certifications: UL1973, UL9540A, UN38.3, IEC62133, EN61000 Expected Cycle Life: ≥ 3,100 cycles at 80 % DOD @25 °C</p>	2 Jobs		

<p>Brand NARADA or equivalent with provision of moveable Rack for batteries</p> <p><b>MS Solar Stand:</b> MS Square Pipe 3" x 3" SWG 16, MS Square Pipe 2" x 2" SWG 16, MS Square Pipe 1.5" x 1.5" SWG 16 with nut bolts and Washers, Painted Complete</p> <p><b>DC Cables, Breakers:</b> 2 core 6mm, 2 core 4mm, 2 core 1.5mm, With All Breakers as per requirement of solar system</p> <p><b>Civil work for solar stand:</b> Complete Civil Work which includes columns and foundation (1.5ft x 1.5ft x upto 3 ft height) for Solar Stand with its Installation as per requirement</p> <p><b>Installation and Transportation:</b> Commissioning, Installation, Testing of Solar System and Transportation Charges.</p> <p><b>Post installation Services:</b> The package includes six months comprehensive service coverage for the solar system, comprising routine maintenance visits, repair support, and replacement of minor wear-and-tear components as required.</p>			
---	--	--	--

- The performance period will be 210 Days / 7 months, which will start from the date of issuing work/supply order.
- The successful bidder(s) is/are required to furnish a performance guarantee, equivalent to 10% of the contract amount within fourteen days of issue of supply/work order/award of contract. Otherwise, equivalent amount will be retained from payment for performance period (6 months). This amount will be treated as compensation for any loss resulting from the contractor's failure to complete its obligation.
- Performance guarantee/security will be forfeited in case of failure to successfully complete the performance period.

## 1. COMPLETION PERIOD

The allotted work should be completed within specified period of 30 days after the receipt of work order. If work is not completed within the allocated time period then the contractor has to pay delay charges @ 0.01% of the total cost of the work for each delayed day.

The bidder will be responsible for complete execution of the work and for its free repair services during performance period.

## 2. MODE OF PAYMENT

No advance payment will be made to any firm against any supply /work order in any case. No such demand will be entertained and failure to execute work against supply order will lead to forfeiture of the earnest money. Payment shall be made in the form of cross-cheque, which shall be issued, after receipt of inspection report of PCRWR officer/ engineer. 10%

payment (retention money) will be returned after defect liability period or completion of performance period (six months).

### **3. PRICES**

Prices charged by the bidder for goods delivered and services performed under the contract shall not exceed from the prices quoted by the supplier/contractor in its bid(s).

### **4. EVALUATION CRITERIA**

The bids would be evaluated based upon their technical soundness with the Terms and Conditions set forth in this document. The lowest quoted bid as a whole meeting technical requirement will be selected.

### **5. DEFECT LIABILITY PERIOD**

The defect liability period is six months, In case of failure; the same will be arranged from the retention money.

The procuring agency may cancel the tender as per PPRA rules. PPRA rules and other prevailing government rules shall be applicable.

The representative of PCRWR, or designated Engineer, may examine any item, equipment, material, supplies and/or services to be provided under this contract, at any time prior to expiry of this contract. He may reject or require alteration thereto or replacements thereof as a condition of acceptance. The findings shall be conclusive except with respect to defects or fraud, which may appear later. Accomplishment and/or execution of each item of work (total or a part thereof) shall be considered subject to the verification by the officer /Engineer/technician designated by the PCRWR.

**GENERAL INFORMATION ABOUT THE CONTRACTOR/FIRM**

- 1- Name of the Firm: \_\_\_\_\_
  
- 2- Main Line or Business: \_\_\_\_\_
  
- 3- Type of Organization: \_\_\_\_\_
  
- 4- Phone No(s) \_\_\_\_\_
  
- 5- Mailing Address: \_\_\_\_\_
  
- 6- Fax No. \_\_\_\_\_
  
- 7- E-mail \_\_\_\_\_
  
- 8- Contact Person: \_\_\_\_\_
  
- 9- Income tax and sales tax numbers-----

**SIGNATURE & SEAL OF CONTRACTOR/FIRM**

**INSTRUCTIONS TO TENDERERS  
(CLAUSE REFERENCE)**

## INSTRUCTIONS TO TENDERERS (CLAUSE REFERENCE)

### 1.1 Name and Address of Employer

Regional Director,  
Pakistan Council of Research in Water Resources  
Main University road, Behind Naseer Tower, Near KW&SB  
Reservoir, Gulistan-e- Johar, Block-1, Karachi  
Phones: (92-21)34028062, 99333717-8  
E-mail: [pcrwr.karachi@gmail.com](mailto:pcrwr.karachi@gmail.com)

### 1.2 Tender shall quote entirely in Pak. Rupees. The payment shall be made to the supplier in Pak. Rupees.

#### **The tenderer will provide the total cost indicating all applicable taxes.**

- a. The tenderer / contractor / manufacturer / distributor / authorized dealer / agents have the financial, technical capability necessary to carry out mentioned works as per mentioned specification as per Bidding Documents.
- b. Tender validity is 210 days, after issuance of Work/Supply order(s) at PCRWR Regional Office, Karachi, at their own expenses and even at short notice if needed. Payment will be made after full satisfaction (demonstration will be mandatory). Delivery of equipment/instruments/completion of work will be in accordance to No. of days specified in the work /supply order.
- c. If supplied items are found sub-standard or in-genuine, the contract will be cancelled and Security Deposits / Performance Security will be forfeited and the Firm will also be blacklisted.
- d. The Firms must have Sales Tax Registration Certificate, National Tax Number, SRB Registration and Vender Number and have proper offices, shops/stores as well as facilities of telephone/fax and email. The Firm must be active filer.
- e. Submission of any false statement/documents including concealing of information is likely to disqualify the bidder.
- f. This Institute reserves the right to accept or reject any or all tenders without assigning any reason in accordance with rule 33(1) of PPRA-2004.

### 1.3 (a) A detailed description of the Works, essential technical and performance characteristic.

(b) Complete set of technical information, description data, literature and drawings as required in accordance with Tendering Data, Specific Works Data. This will include but

not be limited to a sufficient number of drawings, photographs, catalogues, illustrations and such other information as is necessary to illustrate clearly the significant characteristics such as general construction dimensions and other relevant information about the works to be performed.

**1.4 Amount of Tender Security**

5% of the quoted price

**1.5 Period of Tender Validity**

210 days from the date of tender opening.

**1.6 Number of Copies of the Tender to be submitted**

One original plus 1 copy.

**1.7 Deadline for Submission of Tenders**

11:00 AM, 25<sup>th</sup> August, 2025

**1.8 Venue, Time, and Date of Tender Opening**

PCRWR Regional Office,  
Main University Road, Behind Naseer Tower,  
Near KW&SB Reservoir, Gulistan-e-Johar, Block-1, Karachi.

Time: 11:30 AM

Date: 25<sup>th</sup> August, 2025

**1.9 Responsiveness of Tenders**

- (i) the Tender is valid till required period,
- (ii) the Tender prices are firm during currency of contract.
- (iii) completion period offered is within specified limits,
- (iv) the Tenderer/Manufacturer is eligible to Tender and possesses the requisite experience,
- (v) the Tender does not deviate from basic technical requirements and
- (vi) the Tenders are generally in order, etc.

**FORMS OF TENDER  
AND  
SCHEDULES TO TENDER**

**SCHEDULE - TO TENDER**

**FORM OF TENDER**

(LETTER OF OFFER)

To: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Gentlemen,

1. Having examined the Tender Documents including Instructions to Tenderers, Tendering Data, Conditions of Contract, Contract Data, Specification, and Addenda Nos. \_\_\_\_\_ for the “PROCUREMENT OF LABORATORY EQUIPMENT / INSTRUMENTS AND INSTALLTION OF SOLAR SYSTEM UNDER THE PROJECT “STRENGTHENING OF PCRWR WATER QUALITY LABORATORIES IN SINDH IN COLLABORATION WITH UNICEF” We, the undersigned, being a company doing business under the name of and address

\_\_\_\_\_ and being duly incorporated under the laws of Pakistan hereby offer to execute and complete such works and remedy any defects therein in conformity with the said Documents including Addenda thereto for the Total Tender Price of Rupees ( \_\_\_\_\_ ) or such other sum as may be ascertained in accordance with the said Documents.

- 2. We understand that all the Schedules attached hereto form part of this Tender.
- 3. As security for due performance of the undertakings and obligations of this Tender, we submit herewith a Tender Security in the amount of (Rs.) \_\_\_\_\_ drawn in your favor or made payable to you and valid for a period of 28 days beyond the period of validity of Tender.
- 4. We undertake, if our Tender is accepted, to commence the Works and to complete the whole of the Works comprised in the Contract. Comprised in the Contract within the time(s) stated in Contract Date.
- 5. We agree to abide by this Tender for the period of 210 days from the date fixed for receiving the same and it shall remain binding upon us and may be accepted at any time before the expiration of that period.

6. Unless and until a formal Agreement is prepared and executed, this Tender, together with your written acceptance thereof, shall constitute a binding contract between us.

7. We undertake, if our Tender is accepted, to execute the Performance Security referred to in Conditions of Contract for the due performance of the Contract.

8. We understand that you are not bound to accept the lowest or any Tender you may receive.

9. We do hereby declare that the Tender is made without any collusion comparison of figures or arrangement with any other person or persons making a Tender for the Works.

Dated this \_\_\_\_\_ day \_\_\_\_\_

Signature: \_\_\_\_\_

in the capacity of \_\_\_\_\_ duly authorized to sign tenders  
for and on behalf of

---

(Name of Bidder in Block Capitals)  
(Seal)

Address:

---

---

---

Witness:

Signature: \_\_\_\_\_

Name: \_\_\_\_\_

Address. \_\_\_\_\_

\_\_\_\_\_

Occupation \_\_\_\_\_

## FORM OF CONTRACT AGREEMENT

THIS CONTRACT AGREEMENT (hereinafter called the “Agreement”) made on the \_\_\_\_ day of \_\_\_\_\_ between REGIONAL DIRECTOR, PCRWR, KARACHI (hereinafter called the “Employer”) of the one part and \_\_\_\_\_ (hereinafter called the “Contractor”) of the other part.

WHEREAS the Employer is desirous that certain Works, viz \_\_\_\_\_ should be executed by the Contractor and has accepted a Bid by the Contractor for the execution and completion of such Works and the remedying of any defects therein.

NOW this Agreement witness as follows:

1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract hereinafter referred to.
2. The following documents after incorporating addenda, if any except those parts relating to Instructions to Bidders, shall be deemed to form and be read and construed as part of this Agreement, viz:
  - (a) The Letter of Acceptance;
  - (b) The completed Form of Bid along with Schedules to Bid;
  - (c) Conditions of Contract & Contract Data;
  - (d) The priced Schedule of Prices;
3. In consideration of the payments to be made by the Employer to the Contractor as hereinafter mentioned, the Contractor hereby covenants with the Employer to execute and complete the Works and remedy defects therein in conformity and in all respects within the provisions of the Contract.
4. The Employer hereby covenants to pay the Contractor, in consideration of the execution and completion of the Works as per provisions of the Contract, the Contract Price or such other sum as may become payable under the provisions of the Contract at the times and in the manner prescribed by the Contract.

IN WITNESS WHEREOF the parties hereto have caused this Contract Agreement to be executed on the day, month and year first before written in accordance with their respective laws.

Signature of the Contactor

\_\_\_\_\_

(Seal)

Signature of the Employer

\_\_\_\_\_

(Seal)

Signed, Sealed and Delivered in the presence of:

Witness:

\_\_\_\_\_

(Name, Title and Address)

Witness:

\_\_\_\_\_

(Name, Title and Address)

**(INTEGRITY PACT)**

**DECLARATION OF FEES, COMMISSION AND BROKERAGE ETC.  
PAYABLE BY THE SUPPLIERS OF GOODS, SERVICES & WORKS IN  
CONTRACTS WORTH RS. 10.00 MILLION OR MORE**

Contract No. \_\_\_\_\_ Dated \_\_\_\_\_

Contract Value: \_\_\_\_\_

Contract Title: \_\_\_\_\_

..... [name of Supplier/Contractor] hereby declares that it has not obtained or induced the procurement of any contract, right, interest, privilege or other obligation or benefit from Government of Pakistan (GoP) or any administrative subdivision or agency thereof or any other entity owned or controlled by GoP through any corrupt business practice.

Without limiting the generality of the foregoing, I represents and warrants that it has fully declared the brokerage, commission, fees etc. paid or payable to anyone and not given or agreed to give and shall not give or agree to give to anyone within or outside Pakistan either directly or indirectly through any natural or juridical person, including its affiliate, agent, associate, broker, consultant, director, promoter, shareholder, sponsor or subsidiary, any commission, gratification, bribe, finder's fee or kickback, whether described as consultation fee or otherwise, with the object of obtaining or inducing the procurement of a contract, right, interest, privilege or other obligation or benefit in whatsoever form from GoP, except that which has been expressly declared pursuant hereto.

I certify that it has made and will make full disclosure of all agreements and arrangements with all persons in respect of or related to the transaction with GoP and has not taken any action or will not take any action to circumvent the above declaration, representation or warranty.

I accept full responsibility and strict liability for making any false declaration, not making full disclosure, misrepresenting facts or taking any action likely to defeat the purpose of this declaration, representation and warranty. It agrees that any contract, right, interest, privilege or other obligation or benefit obtained or procured as aforesaid shall, without prejudice to any other rights and remedies available to GoP under any law, contract or other instrument, be voidable at the option of GoP.

Notwithstanding any rights and remedies exercised by GoP in this regard, \_\_\_\_\_ [name of Supplier/Contractor] agrees to indemnify GoP for any loss or damage incurred by it on account of its corrupt business practices and further pay compensation to GoP in an amount equivalent to ten time the sum of any commission, gratification, bribe, finder's fee or kickback given by \_\_\_\_\_ [name of Supplier/Contractor] as aforesaid for the purpose of obtaining or inducing the procurement of any contract, right, interest, privilege or other obligation or benefit in whatsoever form from GoP.

Name of Employer.....

Name of Seller/Supplier:/Contractor .....

Signature: .....

Signature: .....

[Seal]

[Seal]